



CERTIFICATE OF CALIBRATION

Model: 4010

This is to certify that all the products distributed by Rubbermaid Commercial Products are calibrated at the time of manufacturing to tolerances in the specification. This does not certify that the products have not been altered by tampering or damaged by misuse or mishandling. This is not meant to supersede individual or internal calibration programs.

Section 1: Calibration Method

Model 4010 is calibrated to exact specification during manufacture in China.

Section 2: Calibration Procedure

Calibration procedure may only be performed by approved operators. Rubbermaid manufacturing procedures require verification of conformance to product specifications by production personnel at the time of manufacture. Product release is contingent on acceptance by Inspection personnel according to Procedure 10-4-01 and is based on sampling plan ANSI/ASQC Standard Z1.4-2003

Section 3: Performance Criteria for 100 lb. Capacity

1% of applied weight at 25° C 10° C to 20° C with less than .05% error

Checkpoint	Tolerance
50 lb.	+/- 0.5 lb.
100 lb.	+/- 1 lb.
150 lb.	+/- 1.5 lb.

Section 4: Environmental Conditions During Testing

Ambient temperature of 77° F ; Relative humidity of 65%

Section 5: Pertinent Information

Weights used for the calibration / verification procedure are certified to tolerances found in ASTM E617

Rubbermaid Commercial Products, LLC Newell Brands Company
8900 North Point Executive Drive
Huntersville, NC 28078

Christian Zhao, Senior quality manager

Signature

June 15, 2023

Date

This document cannot be reproduced without the written consent of Rubbermaid Inc.